

Certified to ISO 9001 and ISO 13485

Quality Mon

Oxygen Monitor Quality Mon O₂



CE 0482

Product Data Sheet Version 1.0

Purpose

Measuring and monitoring of oxygen concentration in the breathing air mixture provided to the patients by an air-oxygen mixer (e.g. Quality Mix).

Specifications	Values
Type of gas	Oxygen
Measuring range	18 to 100% of oxygen in gas mixture
Scaling	0,1
Accuracy	3% relative
Operating temperature	0 to + 45°C
Alarm	visual and audible
Alarm limits	18–101% (manual override 15–101%)
Autoset	±2% of current measured value
Alarm suppression	30 sec.
Dimensions W*D*H	61 x 110 x 92mm
Weight	500g with batteries 3xTyp AAA
O ₂ sensor	OOM111

Made in Germany




DEHAS Medizintechnik & Projektierung GmbH, Wesloer Str. 112, 23568 Lübeck, Germany

HVS Medizintechnik GmbH & Co. KG, Westendstraße 21 • D-87488 Betzigau

Tel.: +49 831 590 97 94 Fax: +49 831 590 97 69

Mail: oliverhoernla@hvshoernla.de • www.hvshoernla.de • www.sauerstoffmischer.com